

10	<div>Ameliyat Masası</div> <div>Operating Table</div> <div>طاولة عمليات</div>
----	---

- (ÖNEMLİ) Dikkatli Okuyun/ (IMPORTANT) Read Carefully;

No	Tedarikçinin Sorumlulukları	Obligations of the supplier	مسؤوليات المورد
1	Teklif verilen cihazların tüm özellikleri detaylı olarak sağlanmalıdır. Aksi takdirde, detaylı olmayan gönderilen herhangi bir katalog, özellik defteri vs. ürünün reddedilmesine neden olabilir.	All specifications of the provided items must be provided in details. As a result, any submitted catalogue with no clear specifications might lead for the rejection of the item.	يجب توضيح جميع مواصفات الأجهزة المقدمة بالتفصيل. نتيجة لذلك ، قد يؤدي أي كتالوج مقدم بدون مواصفات واضحة إلى رفض الجهاز
2	Tedarikçi, sabit elektrik akımı gerektiren herhangi bir cihaz için bir UPS sağlamalıdır.	The supplier has to provide a UPS for any device that requires constant electrical current	يجب على المورد توفير UPS لأي جهاز يتطلب تيارًا كهربائيًا ثابتًا
3	Cihaz, sonuç okumaları, seçenekler ve kontrol için İngilizce dilini desteklemelidir.	The Device must support English language for results readings, options and control.	يجب أن يدعم الجهاز اللغة الإنجليزية لقراءات النتائج والخيارات والتحكم.
4	Tedarikçi cihazı teslim etmeli, kurmalı ve çalıştırmalıdır. Ayrıca tedarikçi, UIYV ekibine gerekli eğitimi vermekle yükümlüdür.	The supplier must deliver, install and operate the device. Also, the supplier is obliged to provide the required training to UIYV team	يجب على المورد تسليم الجهاز وتركيبه وتشغيله. أيضًا ، يلتزم المورد بتوفير التدريب المطلوب لفريق UIYV
5	Cihaz, teslim tarihinden itibaren beş iş günü içinde kurulmalı ve çalıştırılmalıdır. Bu nedenle, kurulumda gerçekleştirilen herhangi bir gecikme olursa, teslimatta bir gecikme olarak sayılacaktır.	The device must be installed and operated within five working days from the delivery date. Therefore, any delay in installation will be considered a delay in delivery.	يجب تركيب الجهاز وتشغيله خلال خمسة أيام عمل من تاريخ التسليم. لذلك ، فإن أي تأخير في التركيب سيُعتبر تأخيرًا في التسليم.
6	Cihaz, kullanıcı dışı herhangi bir hasar için en az iki yıl ücretsiz tamir garantisi edilmelidir. Ayrıca, tedarikçi, iki ay içinde iki onarım arızasından sonra cihazı masrafları tedarikçiye ait olmak üzere değiştirmelidir.	The device must be guaranteed for at least two years free of charge for any non-user damage. Also, The supplier must replace the device on his own cost after two repair failures within two months.	يجب ضمان الجهاز لمدة عامين على الأقل مجانًا لأي ضرر غير مستخدم. أيضًا ، يجب على المورد استبدال الجهاز على نفقته الخاصة بعد فشل إصلاحه لمدة مرتين في غضون شهرين.
7	Tedarikçi, UIYV talebi üzerine teslim tarihinden itibaren en az 5 yıl süreyle cihazın gerekli tüm yedek parça ve aksesuarlarını sağlamakla yükümlüdür.	The supplier is responsible to provide all required spare parts and accessories of the device for at least 5 years from the delivery time upon UIYV request.	المورد مسؤول عن توفير جميع قطع الغيار والملحقات المطلوبة للجهاز لمدة 5 سنوات على الأقل من وقت التسليم بناءً على طلب UIYV.
8	Yukarıda belirtilen yükümlülüklerin herhangi bir şekilde ihlali, tedarikçiyi düşük dereceli performansla maruz bırakacak ve buna göre mali cezalar uygulanacaktır.	Any violation of the above-mentioned obligations will expose the supplier to low rated performance and financial penalties will apply accordingly.	أي انتهاك للالتزامات المذكورة أعلاه سيعرض المورد لأداء منخفض التصنيف وسيتم تطبيق غرامات مالية وفقًا لذلك.

No.	Cihaz Özellikleri Device Specifications مواصفات الجهاز	Uyuşursa (v) işaretleyin Check (v) if conforms تحقق ب (v) في حال المطابقة
1	Table movements/ positions: <ul style="list-style-type: none"> Trendelenburg 25 degree at Least. Reverse Trendelenburg 25 degree at least Lateral Tilt 15 degree at Least. 	
2	Anti-static castors with breaks	
3	Foot switch operated electric tabletop height adjustment.	
4	Table- top frame and column cover should be made of stainless steel or equivalent high-quality material. Table base shall be made of rust proof and acid resisting material.	
5	Divided into four parts at least: <ul style="list-style-type: none"> Head support. Back support. Seat. Legs support. 	
6	Detachable leg support with two parts tilt down with 90 degree approx.	
7	Radiolucent table top (from the head to the end of the pelvic)	
8	Movement angle of back support should be specified by the bidder.	
9	Head support shall enable the adjustment of the head angle, and should be detachable to enable the use of other supports when required.	
10	Stainless steel rails on both sides of the table top to attach the required accessories.	
11	Height adjustment range of the table should be specified by the bidder.	
12	Mattress thickness: 40 mm at least	
13	The Mattress shall withstand the usual disinfection and sterilization procedures, and shall be made of non- absorption, and resistant to the chemicals used materials	
14	Length and width of the table should be specified by the bidder .	
15	Max. load (in Kg) should be specified by the bidder	
16	Table should be provided with the following additional accessories: <ul style="list-style-type: none"> Infusion rod Arm support (pair) Shoulder support (pair) Lateral support Leg support (pair) Goepel knee crutch (pair) Abdomen support Body elevator X-Ray Cassette holder Straps (body, hands and feet) Anesthesia screen. 	
17	Manufactured by a world famous company with CE Mark	
18	The supplier provides maintenance catalogs and electrical and electronic schematics for the table.	