

18	Egzersiz EKG Test Sistemi Exercise ECG Test System نظام ممارسة اختبار تخطيط القلب
----	--

- (ÖNEMLİ) Dikkatli Okuyun/ (IMPORTANT) Read Carefully;

No	Tedarikçinin Sorumlulukları	Obligations of the supplier	مسؤوليات المورد
1	Teklif verilen cihazların tüm özellikleri detaylı olarak sağlanmalıdır. Aksi takdirde, detaylı olmayan gönderilen herhangi bir katalog, özellik defteri vs. ürünün reddedilmesine neden olabilir.	All specifications of the provided items must be provided in details. As a result, any submitted catalogue with no clear specifications might lead for the rejection of the item.	يجب توضيح جميع مواصفات الأجهزة المقدمة بالتفصيل. نتيجة لذلك ، قد يؤدي أي كتالوج مقدم بدون مواصفات واضحة إلى رفض الجهاز
2	Tedarikçi, sabit elektrik akımı gerektiren herhangi bir cihaz için bir UPS sağlamalıdır.	The supplier has to provide a UPS for any device that requires constant electrical current	يجب على المورد توفير UPS لأي جهاز يتطلب تيارًا كهربائيًا ثابتًا
3	Cihaz, sonuç okumaları, seçenekler ve kontrol için İngilizce dilini desteklemelidir.	The Device must support English language for results readings, options and control.	يجب أن يدعم الجهاز اللغة الإنجليزية لقراءات النتائج والخيارات والتحكم.
4	Tedarikçi cihazı teslim etmeli, kurmalı ve çalıştırmalıdır. Ayrıca tedarikçi, UIYV ekibine gerekli eğitimi vermekle yükümlüdür.	The supplier must deliver, install and operate the device. Also, the supplier is obliged to provide the required training to UIYV team	يجب على المورد تسليم الجهاز وتركيبه وتشغيله. أيضًا ، يلتزم المورد بتوفير التدريب المطلوب لفريق UIYV
5	Cihaz, teslim tarihinden itibaren beş iş günü içinde kurulmalı ve çalıştırılmalıdır. Bu nedenle, kurulumda gerçekleştirilen herhangi bir gecikme olursa, teslimatta bir gecikme olarak sayılacaktır.	The device must be installed and operated within five working days from the delivery date. Therefore, any delay in installation will be considered a delay in delivery.	يجب تركيب الجهاز وتشغيله خلال خمسة أيام عمل من تاريخ التسليم. لذلك ، فإن أي تأخير في التركيب سيُعتبر تأخيرًا في التسليم.
6	Cihaz, kullanıcı dışı herhangi bir hasar için en az iki yıl ücretsiz tamir garantisi edilmelidir. Ayrıca, tedarikçi, iki ay içinde iki onarım arızasından sonra cihazı masrafları tedarikçiye ait olmak üzere değiştirmelidir.	The device must be guaranteed for at least two years free of charge for any non-user damage. Also, The supplier must replace the device on his own cost after two repair failures within two months.	يجب ضمان الجهاز لمدة عامين على الأقل مجانًا لأي ضرر غير مستخدم. أيضًا ، يجب على المورد استبدال الجهاز على نفقته الخاصة بعد فشل إصلاحه لمدة مرتين في غضون شهرين.
7	Tedarikçi, UIYV talebi üzerine teslim tarihinden itibaren en az 5 yıl süreyle cihazın gerekli tüm yedek parça ve aksesuarlarını sağlamakla yükümlüdür.	The supplier is responsible to provide all required spare parts and accessories of the device for at least 5 years from the delivery time upon UIYV request.	المورد مسؤول عن توفير جميع قطع الغيار والملحقات المطلوبة للجهاز لمدة 5 سنوات على الأقل من وقت التسليم بناءً على طلب UIYV.
8	Yukarıda belirtilen yükümlülüklerin herhangi bir şekilde ihlali, tedarikçiyi düşük dereceli performansla maruz bırakacak ve buna göre mali cezalar uygulanacaktır.	Any violation of the above-mentioned obligations will expose the supplier to low rated performance and financial penalties will apply accordingly.	أي انتهاك للالتزامات المذكورة أعلاه سيعرض المورد لأداء منخفض التصنيف وسيتم تطبيق غرامات مالية وفقًا لذلك.

No.	Cihaz Özellikleri Device Specifications مواصفات الجهاز	Uyuşursa (v) işaretleyin Check (v) if conforms تحقق ب (v) في حال المطابقة
1	Acquisition Module:	
	1.1 Dimensions (L x W x D): 5.3" x 4.1" x 1.4" (135 mm x 105 mm x 35 mm)	
	1.2 Weight: 4.8 ounces (135 g)	
	1.3 Integrated defibrillation protection	
	1.4 10-lead patient cable	
	1.5 Frequency response: 0.05–150 Hz	
	1.6 Sampling rate: 600 sps	
2	Filter	
	2.1 25/35 Hz muscle tremor filter	
	2.2 Selectable 50/60 Hz	
	2.3 Digital baseline filter	
3	Minimum Computer Requirements	
	3.1 Laptop PC configuration	
	3.2 Processor: i7 7th gen	
	3.3 Ports: 3 USB and 2 serial	
	3.4 Memory: 8 GB	
	3.5 Monitor: 1280 x 1024 true color, 15.6" or higher	
	3.6 Hard disk: At least 3TB	
	3.7 Operating system: Windows 7 64bit I/O: CD-RW and DVD 3.8 Peripherals: Keyboard, mouse and printer color . PC running only CPWS and EHR application during tests	
4	Software	
	4.1 Monitoring formats: 1 x 12, 4 x 3, 2 x 6	
	4.2 Print formats: 3 x 4, 2 x 6, Summary, ST Trends, ST Measurements, Average Complexes and Full Disclosure	
	4.3 Print speeds: 5, 10, 12.5, 25, 50 and 100 mm/sec	
	4.4 Lead sensitivity: 5, 10 and 20 mm/mV	
5	Treadmill	
	5.1 0.5 to 12 mph	
	5.2 0 to 25% elevation	
	5.3 22" x 63" running surface	
	5.4 2 hp motor; 220 volt	
	5.5 Antislip strips	
	5.6 Ultraquiet, long running surface	
	5.7 User capacity: 500 lbs	
	5.8 Floor surface requirements: Level 33" x 88" surface	
6	Bike	
	6.1 Dimensions (L x W x D): 45.3" x 23.6" x 45" (115 cm x 60 cm x 114 cm)	
	6.2 Weight: 123.5 lbs (56 kg)	
	6.3 Maximum weight of rider: 353 lbs (160 kg)	
	6.4 Power requirements: 115/230 VAC, 50/60 Hz (130 VA)	
	6.5 Adjustable seat and handlebars	
7	Power requirements	
	7.1 The item must work on 220V electricity without the need to remove the battery for recharging.	