

19	ELISA Cihazı ELISA Device ELISA جهاز
----	---

- (ÖNEMLİ) Dikkatli Okuyun/ (IMPORTANT) Read Carefully;

No	Tedarikçinin Sorumlulukları	Obligations of the supplier	مسؤوليات المورد
1	Teklif verilen cihazların tüm özellikleri detaylı olarak sağlanmalıdır. Aksi takdirde, detaylı olmayan gönderilen herhangi bir katalog, özellik defteri vs. ürünün reddedilmesine neden olabilir.	All specifications of the provided items must be provided in details. As a result, any submitted catalogue with no clear specifications might lead for the rejection of the item.	يجب توضيح جميع مواصفات الأجهزة المقدمة بالتفصيل. نتيجة لذلك ، قد يؤدي أي كتالوج مقدم بدون مواصفات واضحة إلى رفض الجهاز
2	Tedarikçi, sabit elektrik akımı gerektiren herhangi bir cihaz için bir UPS sağlamalıdır.	The supplier has to provide a UPS for any device that requires constant electrical current	يجب على المورد توفير UPS لأي جهاز يتطلب تيارًا كهربائيًا ثابتًا
3	Cihaz, sonuç okumaları, seçenekler ve kontrol için İngilizce dilini desteklemelidir.	The Device must support English language for results readings, options and control.	يجب أن يدعم الجهاز اللغة الإنجليزية لقراءات النتائج والخيارات والتحكم.
4	Tedarikçi cihazı teslim etmeli, kurmalı ve çalıştırmalıdır. Ayrıca tedarikçi, UIYV ekibine gerekli eğitimi vermekle yükümlüdür.	The supplier must deliver, install and operate the device. Also, the supplier is obliged to provide the required training to UIYV team	يجب على المورد تسليم الجهاز وتركيبه وتشغيله. أيضًا ، يلتزم المورد بتوفير التدريب المطلوب لفريق UIYV
5	Cihaz, teslim tarihinden itibaren beş iş günü içinde kurulmalı ve çalıştırılmalıdır. Bu nedenle, kurulumda gerçekleştirilen herhangi bir gecikme olursa, teslimatta bir gecikme olarak sayılacaktır.	The device must be installed and operated within five working days from the delivery date. Therefore, any delay in installation will be considered a delay in delivery.	يجب تركيب الجهاز وتشغيله خلال خمسة أيام عمل من تاريخ التسليم. لذلك ، فإن أي تأخير في التركيب سيُعتبر تأخيرًا في التسليم.
6	Cihaz, kullanıcı dışı herhangi bir hasar için en az iki yıl ücretsiz tamir garantisi edilmelidir. Ayrıca, tedarikçi, iki ay içinde iki onarım arızasından sonra cihazı masrafları tedarikçiye ait olmak üzere değiştirmelidir.	The device must be guaranteed for at least two years free of charge for any non-user damage. Also, The supplier must replace the device on his own cost after two repair failures within two months.	يجب ضمان الجهاز لمدة عامين على الأقل مجانًا لأي ضرر غير مستخدم. أيضًا ، يجب على المورد استبدال الجهاز على نفقته الخاصة بعد فشل إصلاحه لمدة مرتين في غضون شهرين.
7	Tedarikçi, UIYV talebi üzerine teslim tarihinden itibaren en az 5 yıl süreyle cihazın gerekli tüm yedek parça ve aksesuarlarını sağlamakla yükümlüdür.	The supplier is responsible to provide all required spare parts and accessories of the device for at least 5 years from the delivery time upon UIYV request.	المورد مسؤول عن توفير جميع قطع الغيار والملحقات المطلوبة للجهاز لمدة 5 سنوات على الأقل من وقت التسليم بناءً على طلب UIYV.
8	Yukarıda belirtilen yükümlülüklerin herhangi bir şekilde ihlali, tedarikçiyi düşük dereceli performansla maruz bırakacak ve buna göre mali cezalar uygulanacaktır.	Any violation of the above-mentioned obligations will expose the supplier to low rated performance and financial penalties will apply accordingly.	أي انتهاك للالتزامات المذكورة أعلاه سيعرض المورد لأداء منخفض التصنيف وسيتم تطبيق غرامات مالية وفقًا لذلك.

No.	Cihaz Özellikleri Device Specifications مواصفات الجهاز	Uyuşursa (v) işaretleyin Check (v) if conforms تحقق ب (v) في حال المطابقة
1	Power supply: AC100-240V 50-60Hz.	
2	The device is made of durable materials that do not rust and can be cleaned and sterilized with solutions available in the market	
3	It has a color touch screen of at least 10 inches to display and control parameters	
4	It has a cartridge with a capacity of at least 96 samples	
5	It uses a wavelength ranging from 200-1000 nm	
6	English operation system available for complete edit, measurement, save data	
7	Able to use individually or connect with PC or PAD to control and export data	
8	8 positions optical filter wheel, equipped with 4 standard optical filters and optional filters available	
9	8 channel vertical optical path, zero dispersion single mode fiber measurement system, automatic plate well center position function	
10	96-well visual layout helps to mark blank, sample, positive/negative control, quality control and multi-value control	
11	Multi-choice tests on single plate.	
12	Single or double wavelength measurement	
13	Self-checking optical path, mechanical motion.	
14	Plate shaking function, time and speed are adjustable	
15	Designed with light source energy saving.	
16	Different kind of data export ports available	
17	Measurement results can be exported to "MS Excel" file	
18	Ambient temperature: 4°C~45°C	
19	The relative humidity: ≤70%	
20	Speed: Single wavelength <15s/96well, double wavelength <28s/96 well (Routine mode)	
21	To be of European manufacture or of quality and hold a quality mark	
22	The supplier provides a kit with enough materials to work for more than a month continuously	
23	The supplier offers a computer, either a laptop or a desktop, with a software linking the device with the original copy of the program	
24	The supplier provides all maintenance and operation blocks, a schedule of the necessary spare parts and their price	
25	Equipped with a printer (Bidder must mention the specifications of the printer)	