

7	<b>Tarayıcılar, Ultrasonik Obstetrik / Jinekolojik</b> <b>Scanners, Ultrasonic Obstetric / Gynecologic</b> <b>إيكو نسائية</b>
---	---

**- (ÖNEMLİ) Dikkatli Okuyun/ (IMPORTANT) Read Carefully;**

No	Tedarikçinin Sorumlulukları	Obligations of the supplier	مسؤوليات المورد
1	<b>Teklif verilen cihazların tüm özellikleri detaylı olarak sağlanmalıdır. Aksi takdirde, detaylı olmayan gönderilen herhangi bir katalog, özellik defteri vs. ürünün reddedilmesine neden olabilir.</b>	All specifications of the provided items must be provided in details. As a result, any submitted catalogue with no clear specifications might lead for the rejection of the item.	يجب توضيح جميع مواصفات الأجهزة المقدمة بالتفصيل. نتيجة لذلك ، قد يؤدي أي كتالوج مقدم بدون مواصفات واضحة إلى رفض الجهاز
2	<b>Tedarikçi, sabit elektrik akımı gerektiren herhangi bir cihaz için bir UPS sağlamalıdır.</b>	The supplier has to provide a UPS for any device that requires constant electrical current	يجب على المورد توفير UPS لأي جهاز يتطلب تيارًا كهربائيًا ثابتًا
3	<b>Cihaz, sonuç okumaları, seçenekler ve kontrol için İngilizce dilini desteklemelidir.</b>	The Device must support English language for results readings, options and control.	يجب أن يدعم الجهاز اللغة الإنجليزية لقراءات النتائج والخيارات والتحكم.
4	<b>Tedarikçi cihazı teslim etmeli, kurmalı ve çalıştırmalıdır. Ayrıca tedarikçi, UIYV ekibine gerekli eğitimi vermekle yükümlüdür.</b>	The supplier must deliver, install and operate the device. Also, the supplier is obliged to provide the required training to UIYV team	يجب على المورد تسليم الجهاز وتركيبه وتشغيله. أيضًا ، يلتزم المورد بتوفير التدريب المطلوب لفريق UIYV
5	<b>Cihaz, teslim tarihinden itibaren beş iş günü içinde kurulmalı ve çalıştırılmalıdır. Bu nedenle, kurulumda gerçekleştirilen herhangi bir gecikme olursa, teslimatta bir gecikme olarak sayılacaktır.</b>	The device must be installed and operated within five working days from the delivery date. Therefore, any delay in installation will be considered a delay in delivery.	يجب تركيب الجهاز وتشغيله خلال خمسة أيام عمل من تاريخ التسليم. لذلك ، فإن أي تأخير في التركيب سيُعتبر تأخيرًا في التسليم.
6	<b>Cihaz, kullanıcı dışı herhangi bir hasar için en az iki yıl ücretsiz tamir garantisi edilmelidir. Ayrıca, tedarikçi, iki ay içinde iki onarım arızasından sonra cihazı masrafları tedarikçiye ait olmak üzere değiştirmelidir.</b>	The device must be guaranteed for at least two years free of charge for any non-user damage. Also, The supplier must replace the device on his own cost after two repair failures within two months.	يجب ضمان الجهاز لمدة عامين على الأقل مجانًا لأي ضرر غير مستخدم. أيضًا ، يجب على المورد استبدال الجهاز على نفقته الخاصة بعد فشل إصلاحه لمدة مرتين في غضون شهرين.
7	<b>Tedarikçi, UIYV talebi üzerine teslim tarihinden itibaren en az 5 yıl süreyle cihazın gerekli tüm yedek parça ve aksesuarlarını sağlamakla yükümlüdür.</b>	The supplier is responsible to provide all required spare parts and accessories of the device for at least 5 years from the delivery time upon UIYV request.	المورد مسؤول عن توفير جميع قطع الغيار والملحقات المطلوبة للجهاز لمدة 5 سنوات على الأقل من وقت التسليم بناءً على طلب UIYV.
8	<b>Yukarıda belirtilen yükümlülüklerin herhangi bir şekilde ihlali, tedarikçiyi düşük dereceli performansla maruz bırakacak ve buna göre mali cezalar uygulanacaktır.</b>	Any violation of the above-mentioned obligations will expose the supplier to low rated performance and financial penalties will apply accordingly.	أي انتهاك للالتزامات المذكورة أعلاه سيعرض المورد لأداء منخفض التصنيف وسيتم تطبيق غرامات مالية وفقًا لذلك.

No.	Cihaz Özellikleri Device Specifications مواصفات الجهاز	Uyşursa (v) işaretleyin Check (v) if conforms تحقق ب (v) في حال المطابقة
1	Power supply: 50Hz, 220 volts $\pm 10\%$ And an internal battery of the excellent type available in the market.	
2	<b>Two Scanning Modes:</b> M-Mood & B-Mood using electronic scanning probes.	
3	Displaying types: B\M, B\B, M, B at least	
4	The apparatus should be provided with electronics probes (non-mechanical) as the following: <ul style="list-style-type: none"> <li>Convex (3.5 Mhz)</li> <li>Vaginal probe (6-7Mhz) Its angle 100° at least.</li> <li>When offering probes of multi-frequency), the bidder should specify that.</li> </ul> <b>Detailed Probes specification and resolution should be specified by the bidder.</b>	
5	Gray scale 256 at least.	
6	Controlling possibilities: <ul style="list-style-type: none"> <li>TGC, Freezing, and pan Zoom.</li> <li>Picture reversibility R\L.</li> <li>Power control.</li> <li>AGC (gain control).</li> <li>STC (Sensitive Time Control).</li> <li>Contrast, Brightness.</li> </ul> The bidder should specify these Specifications.	
7	Frame rate 30 frame/sec at least.	
8	LCD ,LED Monitor 14 inch at least.	
9	The device is portable and has a beautiful shape and is made of durable materials that are not rust and knurled	
10	32 T/R channels at least. And The bidder should specify these Specifications.	
11	The apparatus has two transducer connections at least.	
12	Cine Memory of 40 frames at least, bidder should specify the capacity of the memory and the storage time for M-mode.	
13	footswitch for freeze.	
14	The apparatus should be provided with a program to calculate the old of the infant (PBD-AC-HC-FL-CRL-GS-LMP----), weight and the expected date of birth.	
15	<b>Measurements and calculations:</b> Distance, Area, Circumference, volume...	
16	Printer (B\W) of the following specifications: <ul style="list-style-type: none"> <li>Zoom.</li> <li>High accuracy, which specified by the bidder.</li> <li>Ability to division of the picture.</li> </ul>	
17	TEN high quality rolls of printer' papers with each unit.	
18	The bidder should submit 10 echo Gel /5/ liter with each unit.	
19	To be manufactured by a well-known company that has the quality and ISO mark in the medical industries	
20	The bidder should provide a table of prices for spare parts / props of two types - kibadi - battery	
21	The bidder is obligated to provide the following / maintenance and operation catalog - an original copy of the software or a backup hard copy of the driver	